

AUG 29 2005

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130492</u> ✓	
Name <u>John T. Ostoj</u>		ID# <u>4823</u> ✓	Date <u>8/25/05</u> ✓
A	Agency <u>Talkeetna AST</u>		Phone # <u>495-5952</u>
Instrument Location <u>Willow Substation</u>			
B	Alco S/N <u>1301573</u> ✓		Target Value <u>.083</u> ✓
Alco Test Value Average <u>.081</u> ✓		High Pressure <u>1000</u>	<u>809/05/05</u> <u>CBE</u>
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco <u>.083</u> ✓	
Signature <u>[Signature]</u>			
(OVER)			

Under the Alaska Rules of Evidence, I certify that:  
(Do Not Write in the area below)

- (a) I am Chris W. Beheim, Director, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Chris W. Beheim 9-7-05  
Chris W. Beheim Date  
Scientific Director  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130492

AUGUST 25, 2005

OPERATOR'S NAME:

OSTOJ/JOHN/T

OPERATOR'S NUMBER: 4823

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI:

TEST

O.L. #: 1

DEPT/AGENCY: TALO

CASE/REPORT: TEST

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: X301573

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.67 in

ALCO TARGET .082 21:16

BLANK TEST .000 21:17

INTERNAL STANDARD VERIFIED 21:17

ALCO TV 29.67 in .081 21:18

BLANK TEST .000 21:18

SUBJECT SAMPLE .000 21:19

BLANK TEST .000 21:20

ALCO TV 29.67 in .083 21:21

BLANK TEST .000 21:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130492

AUGUST 25, 2005

TIME 21:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.68 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`ab cde fgh i j k l m n o  
p q r s t u v w x y z { | } ~